# Mercy EMS CLINICAL BULLETIN








## Equipment Notice

DEFIBRILLATION PAD CONNECTION OPTIONS


These pads may be from Zoll, Medline, or other approved vendors. They do not provide CPR feedback and are being used as backup pads while our primary pads are backordered.

**Option 1**



Direct Connection to Trunk Cable

**Option 2**



Connection Using Cable Adapter

### Take Home Message:

- Either connection method is acceptable.
- Verify connectors are fully seated before use.
- Pads may be from Zoll, Medline, or other approved vendors.
- Backup pads do not provide CPR feedback.

COMPASSION. EXCELLENCE. COMMUNITY.

MERCY EMS